

March 13, 1969

Dr. Edward W. Webster
Department of Radiology
Massachusetts General Hospital
Boston, Massachusetts

Dear Dr. Webster:

Enclosed is the abstract for the in-vivo dosimetry topic of the radiation dosimetry session or for the radiobiology session as is considered suitable by the Program Committee. The abstract is forwarded as per John Laughlin's recent correspondence with me.

Sincerely,

James G. Kereiakes, Ph.D.
Professor of Radiology
(Radiologic Physics)

JGK/sjk

Enclosure

AUTHOR'S ABSTRACT FORM
SECOND INTERNATIONAL CONFERENCE
ON MEDICAL PHYSICS

August 11-15, 1969

Sheraton Boston Hotel, Boston, Massachusetts

Preparing your abstract:

1. PLEASE READ INSTRUCTIONS BEFORE TYPING.
2. Your abstract will be used for any publication EXACTLY as typed in the box on this form.
3. Your entire abstract, including title, author name(s), text, etc., must fit within the box (approximately 250 words). Refer to sample on the other side of this form.
4. Use a typewriter with a carbon ribbon, if possible, or a reasonably used black ribbon. (A new ribbon may smudge and an old ribbon may be too faint to be photographed.) Draw special symbols or notations in black ink.
5. Do not erase. Your abstract will be used EXACTLY as submitted. All smudges, errors and misspellings will show.
6. Send completed abstract, in duplicate, in a large enough envelope to allow it to be mailed without additional folds. FOLD THE ABSTRACT ONLY AT THE ARROWS. Put a hard cardboard backing in the envelope to prevent damage to the abstract in the mail. Mark the envelope in bold letters — DO NOT BEND OR FOLD.
7. To be considered, the abstracts must arrive at the Conference headquarters by February 28, 1969.

See reverse side for sample form

(Do not type in this space)

BONE MARROW DOSIMETRY IN A CO-60 IRRADIATED TISSUE-EQUIVALENT HUMAN PHANTOM
James G. Kereiakes, Edward B. Silberstein, J. Winston Rogers, and Eugene L. Saenger (University of Cincinnati, Cincinnati, Ohio, U.S.A.)

Whole body and partial body irradiation (either upper half or lower half) is being used in the therapy of certain cancers. This therapy is given as a single dose bilaterally with the xiphoid serving as the radiation field boundary for upper and lower half irradiation exposures. Hematological side effects are observed as a result of these exposures. Correlation of these effects with patient midline dose, integral dose, or average dose is difficult; however, possible correlation with "active" bone marrow dose warrants further study, particularly for the partial body procedures where some "active" bone marrow remains protected during the procedure. The purpose in this study was to determine experimentally "active" bone marrow dose under simulated whole body and partial body cobalt-60 exposure conditions for humans. A tissue-equivalent phantom (Rando) containing a human skeleton was used. Capsules filled with lithium fluoride were judiciously placed in bone cavities as demonstrated by radiographs of each phantom section. The cavities selected were based on locations of "active" bone marrow spaces as indicated by Ellis¹. From the average rad dose and "active" bone marrow weight, the integral dose to "active" bone marrow was calculated. "Active" bone marrow integral doses for lower half-body and upper half-body are 68.9% and exposures 37.0%, respectively, of that determined for whole body exposure under the irradiation exposure conditions given above. The implications of these "active" bone marrow dose determinations in the clinical hematological pattern observed for humans irradiated under similar exposure conditions will be discussed.

¹ Ellis, R.E., Phys. Med. Biol. 5:255, 1961. Supported by DASA Contract No. DA-49-146-XZ-315 and by Public Health Service Training Grant No. 57992-02-68, from the National Center for Radiological Health

Fc
O:
on this
line

4 → Fold
ONLY
on this
line

BONE MARROW DOSE IN A ^{60}Co IRRADIATED TISSUE-EQUIVALENT
HUMAN PHANTOM USING LITHIUM FLUORIDE DOSIMETRY

James G. Kerefakes and Eugene L. Saenger

Whole body and partial body irradiation (either upper half or lower half) is being used in the therapy of certain cancers. This therapy is given as a single dose bilaterally with the xiphoid serving as the radiation field boundary for upper and lower half irradiation exposures. Hematological side effects are observed as a result of these exposures. Correlation of these effects to patient integral dose is difficult, however, possible correlation with "active" bone marrow dose warrants further study, particularly for the partial body procedures, where some "active" bone marrow remains protected during the procedure. The purpose in this study was to experimentally determine "active" bone marrow dose under simulated whole body and partial body cobalt-60 exposure conditions for humans. A tissue-equivalent phantom (R&D) containing a human skeleton was used. Capsules filled with lithium fluoride were judiciously placed in bone cavities as demonstrated by radiographs of each phantom section. The cavities selected were based on locations of "active" bone marrow spaces as indicated in the review article by Ellis (). From the average rad dose and "active" bone marrow weight, the integral dose to "active" bone marrow was calculated for each anatomical phantom section. Total gram rads to the whole body "active" bone marrow were found to be 56477, 38903 and 20916 gram.rads for whole body, lower half-body, and upper half-body, respectively, per .R trunk midline air exposure. "Active" bone marrow integral dose for lower half-body exposure is 68.9% of that determined for whole body exposure under the irradiation exposure conditions given above. The importance of these "active" bone marrow dose determinations in the clinical hematological pattern observed for humans irradiated under similar exposure conditions will be discussed.